

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000922  
14:47 07/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:30  
STERILIZING 4:35  
136.5 C 224kPa 4:35  
136.1 C 222kPa 5:05  
136.5 C 224kPa 5:35  
136.3 C 224kPa 6:05  
136.4 C 225kPa 6:35  
136.4 C 224kPa 7:05  
136.4 C 225kPa 7:35  
136.4 C 225kPa 8:05  
136.4 C 225kPa 8:05  
VENTING 8:06  
AIR DRYING 9:01  
CYCLE COMPLETE 9:11

-----  
Digital Signature #  
32864C0967E638D2  
-----