
Midmark M11 - v1.0.5
Total Cycles: 1011

Name: RITTER2
Sterilizer ID: V2313228
04/20/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.4 F 0.0
2:00 201.6 F 2.3
4:00 217.5 F 8.7
6:00 231.9 F 9.4
8:00 238.5 F 12.1
10:00 239.0 F 10.8
12:00 239.4 F 10.2
14:00 253.8 F 17.7
16:00 266.0 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.4 F 29.7
2:00 271.7 F 28.8
3:00 272.4 F 29.7
4:00 272.2 F 29.1

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 213.7 F 0.0
10:00 221.5 F 0.0
15:00 230.1 F 0.0
20:00 227.6 F 0.0
25:00 223.5 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 17:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:56:60
CYCLE COMPLETE

04/20/2022 10:59 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____