
Midmark M11 - v1.0.5
Total Cycles: 858

Name: RITTER2
Sterilizer ID: V2313228
03/02/2022 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 187.0 F 3.9
4:00 197.0 F 9.1
6:00 217.9 F 7.6
8:00 227.4 F 10.2
10:00 232.3 F 8.0
12:00 241.4 F 11.7
14:00 245.1 F 13.5
16:00 258.1 F 20.5
18:00 269.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 30.0
2:00 271.6 F 29.3
3:00 271.9 F 30.0
4:00 272.5 F 29.8

Min 271.3 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 201.6 F 0.0
10:00 214.0 F 0.0
15:00 211.8 F 0.0
20:00 209.4 F 0.0
25:00 209.3 F 0.0
30:00 208.7 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

03/02/2022 01:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____