
Midmark M11 - v1.0.5
Total Cycles: 3615

Name: RITTER2
Sterilizer ID: V2313228
05/14/2024 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 167.4 F 3.2
4:00 192.2 F 8.2
6:00 211.1 F 9.3
8:00 230.5 F 9.6
10:00 232.7 F 8.3
12:00 242.2 F 12.6
14:00 244.5 F 12.8
16:00 257.5 F 19.9
18:00 269.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.6
1:00 272.2 F 29.5
2:00 273.1 F 29.8
3:00 272.3 F 30.1
4:00 272.5 F 30.2

Min 271.7 F 29.5
Max 273.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 232.8 F 0.0
10:00 189.2 F 0.0
15:00 183.9 F 0.0
20:00 178.6 F 0.0
25:00 176.9 F 0.0
30:00 175.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

05/14/2024 12:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____