
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000082
9:55 21/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:42
136.0 C 221kPa 4:42
136.4 C 225kPa 5:12
136.3 C 224kPa 5:42
136.4 C 226kPa 6:12
136.4 C 224kPa 6:42
136.3 C 224kPa 7:12
136.4 C 222kPa 7:42
136.4 C 225kPa 8:12
136.4 C 225kPa 8:12
VENTING 8:12
AIR DRYING 9:05
CYCLE COMPLETE 10:45

Digital Signature #
632DFE682338684F
