
Midmark M11 - v1.0.0
Total Cycles: 728

Name: RITTER 1
Sterilizer ID: V2178198
03/18/2021 06:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 74.2 F 0.0
4:00 80.6 F 0.0
6:00 101.1 F 0.1
8:00 144.8 F 0.2
10:00 185.7 F 0.3
12:00 204.0 F 3.8
14:00 215.4 F 6.8
16:00 224.5 F 10.2
18:00 231.9 F 9.6
20:00 236.8 F 10.4
22:00 233.7 F 7.9
24:00 243.9 F 11.9
26:00 245.1 F 12.8
28:00 257.8 F 19.4
30:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.6
2:00 271.7 F 28.4
3:00 272.2 F 28.5
4:00 271.6 F 28.3

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 215.7 F 0.0
10:00 225.7 F 0.0
15:00 231.8 F 0.0
20:00 227.9 F 0.0
25:00 232.0 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 31:07
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

03/18/2021 07:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____