
Midmark M11 - v1.0.0
Total Cycles: 1596

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2022 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.6 F 0.0
2:00 194.2 F 3.9
4:00 200.3 F 8.4
6:00 216.6 F 8.2
8:00 227.7 F 8.8
10:00 236.7 F 10.2
12:00 241.6 F 12.0
14:00 240.5 F 10.5
16:00 253.8 F 17.3
18:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.8 F 28.4
3:00 272.2 F 29.0
4:00 272.7 F 29.0

Min 271.4 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 232.0 F 0.0
10:00 227.9 F 0.0
15:00 224.9 F 0.0
20:00 226.3 F 0.0
25:00 228.1 F 0.0
30:00 224.6 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

02/03/2022 11:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____