

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2617

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2022 09:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.0 F 0.0  
2:00 86.6 F 0.0  
4:00 102.8 F 0.0  
6:00 125.7 F 0.4  
8:00 157.6 F 3.0  
10:00 185.1 F 6.4  
12:00 215.7 F 10.1  
14:00 222.0 F 8.8  
16:00 226.9 F 7.2  
18:00 236.9 F 11.3  
20:00 231.9 F 6.7  
22:00 242.2 F 12.3  
24:00 239.4 F 10.1  
26:00 252.0 F 16.3  
28:00 262.9 F 22.7  
30:00 272.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.6 F 28.2  
3:00 272.1 F 29.0  
4:00 272.3 F 28.8

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 218.6 F 0.0  
10:00 221.5 F 0.0  
15:00 229.1 F 0.0  
20:00 235.3 F 0.0  
25:00 216.0 F 0.0  
30:00 213.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:18  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

11/29/2022 10:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_