
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000964
16:21 07/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:49
STERILIZING 3:59
136.0 C 227kPa 3:59
136.3 C 233kPa 4:29
136.4 C 232kPa 4:59
136.1 C 230kPa 5:29
136.4 C 231kPa 5:59
136.4 C 233kPa 6:29
136.4 C 231kPa 6:59
136.4 C 233kPa 7:29
136.4 C 233kPa 7:29
VENTING 7:29
AIR DRYING 8:22
CYCLE COMPLETE 11:14

Digital Signature #
BEBCB96A4CF2AEAE
