
Midmark M11 - v1.0.0
Total Cycles: 1973

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2022 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.0 F 0.0
2:00 140.8 F 2.5
4:00 167.0 F 5.9
6:00 195.0 F 10.1
8:00 207.7 F 6.0
10:00 219.0 F 8.0
12:00 234.3 F 12.7
14:00 232.9 F 8.7
16:00 242.0 F 11.0
18:00 243.5 F 12.1
20:00 255.5 F 18.3
22:00 266.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.3 F 28.8
2:00 272.3 F 29.2
3:00 271.4 F 28.7
4:00 272.2 F 28.9

Min 271.3 F 28.5
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 210.4 F 0.0
10:00 203.6 F 0.0
15:00 211.3 F 0.0
20:00 230.1 F 0.0
25:00 231.6 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

05/30/2022 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____