

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002089  
7:44 24/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:37  
PRESSURIZING 3:38  
STERILIZING 4:45  
136.0 C 220kPa 4:45  
136.3 C 220kPa 5:15  
136.2 C 223kPa 5:45  
136.3 C 223kPa 6:15  
136.4 C 224kPa 6:45  
136.4 C 223kPa 7:15  
136.4 C 222kPa 7:45  
136.3 C 220kPa 8:15  
136.3 C 220kPa 8:15  
VENTING 8:15  
AIR DRYING 9:09  
CYCLE COMPLETE 39:09

-----  
Digital Signature #  
4A9384502E1235ED  
-----