
Midmark M11 - v1.0.0
Total Cycles: 5400

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2025 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.1 F 0.0
2:00 138.5 F 2.6
4:00 157.0 F 6.3
6:00 188.9 F 10.2
8:00 216.4 F 4.2
10:00 221.8 F 7.2
12:00 233.6 F 11.6
14:00 231.0 F 7.6
16:00 244.3 F 13.2
18:00 241.5 F 11.4
20:00 254.2 F 17.8
22:00 265.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.2 F 29.1
2:00 271.0 F 28.6
3:00 272.3 F 29.1
4:00 272.6 F 29.2

Min 271.0 F 28.4
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 225.7 F 0.0
10:00 209.3 F 0.0
15:00 197.5 F 0.0
20:00 193.8 F 0.0
25:00 189.6 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:46
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

03/07/2025 04:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____