
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000462
10:19 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:49
STERILIZING 4:03
136.0 C 228kPa 4:03
136.2 C 230kPa 4:33
136.4 C 233kPa 5:03
136.2 C 229kPa 5:33
136.4 C 233kPa 6:03
136.4 C 233kPa 6:33
136.4 C 233kPa 7:03
136.2 C 231kPa 7:33
136.2 C 231kPa 7:33
VENTING 7:33
AIR DRYING 8:28
CYCLE COMPLETE 9:00

Digital Signature #
2A7D3ACB8E399FE7
