
Midmark M11 - v1.0.0
Total Cycles: 2091

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2022 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.1 F 0.0
2:00 135.8 F 0.4
4:00 157.8 F 2.8
6:00 187.7 F 6.0
8:00 205.5 F 10.3
10:00 216.8 F 5.2
12:00 228.1 F 8.4
14:00 238.7 F 12.9
16:00 236.7 F 9.7
18:00 239.7 F 9.2
20:00 247.4 F 14.0
22:00 259.2 F 20.4
24:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.9 F 28.4
3:00 272.2 F 29.0
4:00 272.6 F 28.9

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 199.9 F 0.0
10:00 210.8 F 0.0
15:00 210.5 F 0.0
20:00 211.1 F 0.0
25:00 212.4 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

07/05/2022 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____