
Midmark M11 - v1.0.5
Total Cycles: 2102

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 02:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 190.9 F 4.0
4:00 211.4 F 8.8
6:00 221.6 F 8.1
8:00 228.2 F 9.4
10:00 232.6 F 8.9
12:00 243.2 F 13.3
14:00 243.2 F 12.6
16:00 256.6 F 19.7
18:00 268.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.1 F 30.0
2:00 272.6 F 30.0
3:00 272.6 F 29.7
4:00 272.1 F 29.4

Min 271.2 F 29.0
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 226.5 F 0.0
10:00 247.0 F 0.0
15:00 236.2 F 0.0
20:00 228.7 F 0.0
25:00 228.2 F 0.0
30:00 229.5 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

02/28/2023 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____