
Midmark M11 - v1.0.5
Total Cycles: 950

Name: RITTER2
Sterilizer ID: V2313228
04/04/2022 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 165.7 F 0.1
4:00 180.0 F 0.5
6:00 204.4 F 4.2
8:00 218.7 F 9.2
10:00 229.6 F 9.8
12:00 237.5 F 11.5
14:00 237.4 F 10.0
16:00 237.1 F 9.1
18:00 251.7 F 16.5
20:00 263.9 F 23.8
22:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.7
2:00 272.3 F 29.0
3:00 271.6 F 29.0
4:00 272.6 F 29.7

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 197.9 F 0.0
10:00 214.0 F 0.0
15:00 218.9 F 0.0
20:00 222.3 F 0.0
25:00 224.0 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

04/04/2022 11:36 AM
ACCEPTED: X
REJECTED:
Operator: U02