
Midmark M11 - v1.0.5
Total Cycles: 180

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 151.1 F 2.4
4:00 175.5 F 6.1
6:00 200.9 F 11.1
8:00 215.9 F 5.2
10:00 229.9 F 10.6
12:00 230.7 F 7.6
14:00 242.7 F 12.2
16:00 244.1 F 12.8
18:00 256.9 F 19.6
20:00 268.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.5
2:00 271.7 F 29.5
3:00 271.9 F 29.2
4:00 271.7 F 29.6

Min 271.1 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.0 F 0.0
5:00 216.8 F 0.0
10:00 227.7 F 0.0
15:00 195.9 F 0.0
20:00 187.0 F 0.0
25:00 179.0 F 0.0
30:00 177.1 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:12
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

07/08/2021 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U02