

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000948  
16:38 23/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:38  
PRESSURIZING 3:43  
STERILIZING 4:59  
136.4 C 232kPa 4:59  
136.3 C 232kPa 5:29  
136.4 C 232kPa 5:59  
136.4 C 233kPa 6:29  
136.2 C 231kPa 6:59  
136.4 C 232kPa 7:29  
136.4 C 233kPa 7:59  
136.2 C 231kPa 8:29  
136.2 C 231kPa 8:29  
VENTING 8:30  
AIR DRYING 9:23  
CYCLE COMPLETE 14:43  
-----

Digital Signature #  
2BDF7D71267CA610  
-----