
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000272
10:22 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:15
STERILIZING 4:48
136.0 C 224kPa 4:48
136.3 C 233kPa 5:18
136.4 C 232kPa 5:48
136.4 C 233kPa 6:18
136.4 C 233kPa 6:48
136.4 C 231kPa 7:18
136.4 C 234kPa 7:48
136.4 C 233kPa 8:18
136.4 C 233kPa 8:18
VENTING 8:18
AIR DRYING 9:21
CYCLE COMPLETE 23:36

Digital Signature #
4B3201F8931D3D89
