
Midmark M11 - v1.0.5
Total Cycles: 2275

Name: RITTER2
Sterilizer ID: V2313228
04/20/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 159.7 F 3.1
4:00 176.4 F 7.1
6:00 200.5 F 12.3
8:00 219.9 F 6.6
10:00 232.8 F 12.0
12:00 234.3 F 9.5
14:00 236.8 F 9.2
16:00 249.1 F 15.1
18:00 261.3 F 22.2
20:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 271.7 F 28.9
2:00 271.8 F 29.0
3:00 272.0 F 29.3
4:00 272.1 F 29.6

Min 271.5 F 28.9
Max 272.3 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 207.7 F 0.0
10:00 230.1 F 0.0
15:00 222.9 F 0.0
20:00 218.6 F 0.0
25:00 218.9 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

04/20/2023 11:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____