
Midmark M11 - v1.0.0
Total Cycles: 2559

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.1 F 0.0
2:00 193.1 F 3.4
4:00 195.1 F 8.6
6:00 216.7 F 8.4
8:00 223.8 F 9.0
10:00 234.8 F 9.7
12:00 240.4 F 12.0
14:00 239.9 F 10.6
16:00 254.1 F 17.7
18:00 266.3 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.1
2:00 271.6 F 28.5
3:00 272.7 F 29.3
4:00 271.9 F 28.6

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 227.6 F 0.0
10:00 227.8 F 0.0
15:00 216.6 F 0.0
20:00 226.8 F 0.0
25:00 231.5 F 0.0
30:00 229.1 F 0.0

Drying Complete

FILLING: 2:17
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:00
CYCLE COMPLETE

11/15/2022 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____