
Midmark M11 - v1.0.5
Total Cycles: 4273

Name: RITTER2
Sterilizer ID: V2313228
01/17/2025 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 69.7 F 0.0
4:00 76.0 F 0.0
6:00 100.1 F 0.0
8:00 137.3 F 0.2
10:00 168.8 F 0.5
12:00 197.1 F 2.8
14:00 210.9 F 6.1
16:00 224.0 F 9.7
18:00 233.5 F 12.2
20:00 232.3 F 9.1
22:00 240.2 F 11.2
24:00 242.1 F 12.0
26:00 240.0 F 10.7
28:00 252.9 F 17.1
30:00 263.7 F 23.5
32:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.7 F 28.9
3:00 272.7 F 29.3
4:00 271.6 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 212.4 F 0.0
10:00 192.1 F 0.0
15:00 188.3 F 0.0
20:00 184.2 F 0.0
25:00 185.2 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:11
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:25
CYCLE COMPLETE

01/17/2025 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____