
Midmark M11 - v1.0.5
Total Cycles: 2986

Name: RITTER2
Sterilizer ID: V2313228
10/10/2023 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 164.6 F 2.0
4:00 169.9 F 5.5
6:00 192.4 F 10.1
8:00 218.7 F 5.5
10:00 229.5 F 9.7
12:00 233.1 F 8.9
14:00 242.3 F 12.4
16:00 240.5 F 11.0
18:00 253.3 F 17.4
20:00 264.9 F 24.4
22:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.5
2:00 272.8 F 30.0
3:00 271.6 F 29.2
4:00 272.6 F 29.8

Min 271.5 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 193.1 F 0.0
10:00 192.3 F 0.0
15:00 190.6 F 0.0
20:00 189.3 F 0.0
25:00 187.7 F 0.0
30:00 185.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

10/10/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____