
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000440
11:42 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:48
STERILIZING 3:50
136.4 C 232kPa 3:50
136.3 C 232kPa 4:20
136.4 C 229kPa 4:50
136.2 C 231kPa 5:20
136.2 C 231kPa 5:50
136.4 C 233kPa 6:20
136.4 C 232kPa 6:50
136.4 C 233kPa 7:20
136.4 C 233kPa 7:20
VENTING 7:20
AIR DRYING 8:12
CYCLE COMPLETE 8:21

Digital Signature #
5AF57E4865C1FB27
