
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000897
14:24 02/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:24
STERILIZING 4:32
136.0 C 227kPa 4:32
136.1 C 230kPa 5:02
136.4 C 233kPa 5:32
136.4 C 232kPa 6:02
136.4 C 233kPa 6:32
136.2 C 230kPa 7:02
136.4 C 232kPa 7:32
136.4 C 233kPa 8:02
136.4 C 233kPa 8:02
VENTING 8:02
AIR DRYING 8:57
CYCLE COMPLETE 23:57

Digital Signature #
41A6A8BDB252875F
