
Midmark M11 - v1.0.0
Total Cycles: 3816

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 82.9 F 0.0
4:00 92.2 F 0.0
6:00 119.1 F 0.0
8:00 148.9 F 0.2
10:00 177.0 F 2.7
12:00 203.8 F 5.4
14:00 214.2 F 8.9
16:00 225.4 F 12.4
18:00 226.4 F 6.3
20:00 236.6 F 10.4
22:00 236.3 F 8.5
24:00 243.7 F 12.1
26:00 240.4 F 10.2
28:00 253.2 F 16.5
30:00 264.0 F 22.9
32:00 271.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.8
1:00 272.5 F 28.3
2:00 272.0 F 28.4
3:00 271.9 F 28.0
4:00 272.5 F 28.6

Min 271.4 F 27.7
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 225.9 F 0.0
10:00 208.4 F 0.0
15:00 224.9 F 0.0
20:00 227.6 F 0.0
25:00 228.8 F 0.0
30:00 228.5 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:05
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

09/26/2023 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____