
Midmark M11 - v1.0.0
Total Cycles: 3087

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.3 F 0.0
2:00 86.7 F 0.0
4:00 107.0 F 0.0
6:00 135.3 F 0.5
8:00 166.7 F 2.3
10:00 185.8 F 5.1
12:00 203.2 F 8.5
14:00 220.9 F 12.2
16:00 223.4 F 5.3
18:00 231.8 F 9.3
20:00 241.6 F 13.0
22:00 238.4 F 10.3
24:00 237.4 F 8.3
26:00 247.5 F 14.0
28:00 259.0 F 20.2
30:00 269.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.3 F 28.8
2:00 271.9 F 28.8
3:00 272.0 F 28.6
4:00 272.0 F 28.9

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 220.3 F 0.0
10:00 230.1 F 0.0
15:00 229.9 F 0.0
20:00 230.0 F 0.0
25:00 230.3 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:04
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

04/11/2023 08:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____