
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006838
14:55 17/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 3:00
136.4 C 220kPa 3:00
136.4 C 222kPa 3:30
136.3 C 223kPa 4:00
136.1 C 222kPa 4:30
136.5 C 224kPa 5:00
136.4 C 224kPa 5:30
136.4 C 224kPa 6:00
136.2 C 223kPa 6:30
136.2 C 223kPa 6:30
VENTING 6:30
CYCLE ABORTED 7:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:03

Digital Signature #
3B32280F83537B67
