

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3528

Name: RITTER 1  
Sterilizer ID: V2178198  
07/14/2023 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.5 F 0.0  
2:00 174.0 F 1.4  
4:00 182.4 F 3.9  
6:00 202.8 F 8.0  
8:00 217.4 F 12.3  
10:00 225.4 F 7.0  
12:00 236.9 F 11.9  
14:00 236.4 F 9.2  
16:00 239.0 F 9.8  
18:00 248.0 F 14.1  
20:00 259.7 F 20.5  
22:00 270.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.4 F 28.8  
2:00 271.5 F 28.1  
3:00 271.8 F 28.7  
4:00 272.2 F 28.7

Min 271.1 F 28.0  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 199.7 F 0.0  
10:00 201.3 F 0.0  
15:00 198.6 F 0.0  
20:00 195.3 F 0.0  
25:00 195.2 F 0.0  
30:00 193.7 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 22:52  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

07/14/2023 11:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_