

-----  
Midmark M11 - v1.0.5  
Total Cycles: 618

Name: RITTER2  
Sterilizer ID: V2313228  
12/09/2021 09:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.6 F 0.0  
2:00 109.7 F 0.7  
4:00 114.1 F 2.5  
6:00 142.0 F 6.1  
8:00 172.5 F 10.5  
10:00 206.0 F 3.3  
12:00 220.3 F 7.8  
14:00 232.2 F 12.6  
16:00 235.2 F 9.6  
18:00 237.4 F 9.4  
20:00 247.3 F 14.4  
22:00 259.4 F 21.1  
24:00 270.4 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.2 F 29.3  
2:00 272.1 F 29.7  
3:00 271.8 F 29.0  
4:00 272.5 F 29.8

Min 271.5 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
5:00 231.1 F 0.0  
10:00 230.9 F 0.0  
15:00 233.6 F 0.0  
20:00 226.4 F 0.0  
25:00 219.4 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:46  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:02:52  
CYCLE COMPLETE

12/09/2021 10:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02