
Midmark M11 - v1.0.0
Total Cycles: 3161

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.9 F 0.0
2:00 196.4 F 0.0
4:00 213.0 F 2.0
6:00 227.4 F 7.0
8:00 239.4 F 12.2
10:00 239.0 F 10.0
12:00 237.7 F 8.5
14:00 240.2 F 9.7
16:00 251.6 F 15.7
18:00 264.5 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.5
2:00 271.5 F 27.9
3:00 272.6 F 28.6
4:00 271.7 F 27.9

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 226.4 F 0.0
10:00 231.2 F 0.0
15:00 230.4 F 0.0
20:00 228.3 F 0.0
25:00 227.2 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

04/25/2023 12:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____