

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1812

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2022 12:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.7 F 0.0  
2:00 160.1 F 3.1  
4:00 172.0 F 7.7  
6:00 200.5 F 13.2  
8:00 219.7 F 7.6  
10:00 234.8 F 13.3  
12:00 238.3 F 11.2  
14:00 238.8 F 10.3  
16:00 253.3 F 17.6  
18:00 265.8 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.2  
1:00 273.0 F 29.9  
2:00 272.0 F 29.1  
3:00 273.2 F 29.9  
4:00 272.3 F 29.1  
  
Min 271.9 F 29.0  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
5:00 203.6 F 0.0  
10:00 210.3 F 0.0  
15:00 214.8 F 0.0  
20:00 217.9 F 0.0  
25:00 219.1 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

12/06/2022 01:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_