
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006030
13:12 21/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:51
STERILIZING 3:57
136.0 C 220kPa 3:57
135.8 C 220kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
136.4 C 224kPa 5:57
136.4 C 223kPa 6:27
136.1 C 222kPa 6:57
136.2 C 221kPa 7:27
136.2 C 221kPa 7:27
VENTING 7:27
AIR DRYING 8:17
CYCLE COMPLETE 38:17

Digital Signature #
C8A694BA94541CE8
