
Midmark M11 - v1.0.0
Total Cycles: 2066

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.3 F 0.0
2:00 96.5 F 0.0
4:00 111.7 F 0.9
6:00 141.2 F 2.9
8:00 173.5 F 6.0
10:00 194.5 F 9.6
12:00 209.3 F 13.5
14:00 219.6 F 5.8
16:00 231.4 F 9.7
18:00 237.1 F 10.2
20:00 239.7 F 10.7
22:00 234.9 F 7.7
24:00 248.8 F 14.4
26:00 259.9 F 20.6
28:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.8
2:00 271.3 F 27.9
3:00 272.4 F 28.9
4:00 272.2 F 28.3

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 225.4 F 0.0
10:00 217.9 F 0.0
15:00 222.7 F 0.0
20:00 229.3 F 0.0
25:00 227.5 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 28:54
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:02
CYCLE COMPLETE

06/28/2022 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____