
Midmark M11 - v1.0.0
Total Cycles: 4087

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2023 05:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.4 F 0.0
2:00 128.6 F 1.3
4:00 135.3 F 4.2
6:00 163.2 F 7.7
8:00 186.6 F 12.1
10:00 211.4 F 4.5
12:00 223.1 F 8.7
14:00 234.0 F 13.2
16:00 235.2 F 9.5
18:00 238.7 F 9.8
20:00 245.8 F 13.3
22:00 257.7 F 19.6
24:00 268.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.2 F 29.1
2:00 272.1 F 28.6
3:00 271.7 F 28.7
4:00 272.7 F 29.0

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 216.0 F 0.0
10:00 216.5 F 0.0
15:00 225.3 F 0.0
20:00 226.3 F 0.0
25:00 228.0 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:44
CYCLE COMPLETE

12/05/2023 06:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____