

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000233  
13:40 28/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:31  
PRESSURIZING 3:38  
STERILIZING 5:06  
136.0 C 229kPa 5:06  
136.2 C 232kPa 5:36  
136.2 C 231kPa 6:06  
136.4 C 232kPa 6:36  
136.4 C 233kPa 7:06  
136.3 C 231kPa 7:36  
136.2 C 231kPa 8:06  
136.4 C 233kPa 8:36  
136.4 C 233kPa 8:36  
VENTING 8:36  
AIR DRYING 9:37  
CYCLE COMPLETE 24:37  
-----

Digital Signature #  
B2815D98CE9B4DEF  
-----