
Midmark M11 - v1.0.5
Total Cycles: 3883

Name: RITTER2
Sterilizer ID: V2313228
08/22/2024 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.0 F 0.0
2:00 105.8 F 0.0
4:00 107.7 F 0.1
6:00 134.9 F 1.7
8:00 158.4 F 4.8
10:00 176.2 F 8.2
12:00 192.8 F 11.7
14:00 213.9 F 5.0
16:00 223.8 F 9.1
18:00 235.4 F 13.4
20:00 236.5 F 10.5
22:00 235.3 F 8.6
24:00 247.7 F 14.6
26:00 258.6 F 20.6
28:00 268.5 F 27.0
30:00 271.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 271.9 F 29.7
2:00 271.8 F 29.7
3:00 271.7 F 29.7
4:00 271.8 F 29.6

Min 271.7 F 29.5
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 211.6 F 0.0
10:00 217.0 F 0.0
15:00 201.5 F 0.0
20:00 200.2 F 0.0
25:00 202.7 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:02
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:08:36
CYCLE COMPLETE

08/22/2024 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____