
Midmark M11 - v1.0.0
Total Cycles: 996

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2021 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.1 F 0.0
2:00 142.2 F 1.3
4:00 157.9 F 3.5
6:00 181.8 F 7.0
8:00 201.6 F 11.5
10:00 218.8 F 5.6
12:00 229.4 F 10.5
14:00 231.7 F 7.6
16:00 241.9 F 12.4
18:00 239.8 F 10.4
20:00 252.4 F 16.5
22:00 263.2 F 22.8
24:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.8 F 28.8
3:00 272.2 F 28.8
4:00 271.8 F 28.8

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.9 F 0.0
5:00 216.0 F 0.0
10:00 237.3 F 0.0
15:00 233.2 F 0.0
20:00 227.6 F 0.0
25:00 228.5 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

07/12/2021 01:42 PM
ACCEPTED: X
REJECTED:
Operator: U01