

-----  
Midmark M11 - v1.0.5  
Total Cycles: 677

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2022 01:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.6 F 0.0  
2:00 170.4 F 3.2  
4:00 180.9 F 7.3  
6:00 202.5 F 12.2  
8:00 220.4 F 7.2  
10:00 234.7 F 12.4  
12:00 235.6 F 10.3  
14:00 235.7 F 8.7  
16:00 249.8 F 15.8  
18:00 261.2 F 22.4  
20:00 271.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.3 F 29.9  
2:00 271.5 F 29.2  
3:00 271.6 F 29.9  
4:00 272.2 F 29.7  
  
Min 271.1 F 29.1  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 217.2 F 0.0  
10:00 224.0 F 0.0  
15:00 219.6 F 0.0  
20:00 215.6 F 0.0  
25:00 216.3 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:51  
CYCLE COMPLETE

01/10/2022 02:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_