
Midmark M11 - v1.0.5
Total Cycles: 2266

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.7 F 0.0
2:00 176.6 F 3.2
4:00 197.0 F 8.1
6:00 217.6 F 11.5
8:00 229.5 F 9.1
10:00 235.4 F 9.6
12:00 244.8 F 13.2
14:00 245.1 F 13.1
16:00 258.1 F 20.2
18:00 269.8 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.2
2:00 271.7 F 29.2
3:00 272.1 F 29.0
4:00 271.6 F 29.2

Min 271.3 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 246.7 F 0.0
10:00 223.1 F 0.0
15:00 215.6 F 0.0
20:00 217.4 F 0.0
25:00 222.0 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

04/19/2023 10:43 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____