
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007562
7:26 18/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:50
STERILIZING 5:01
136.4 C 221kPa 5:01
136.2 C 221kPa 5:31
136.0 C 219kPa 6:01
136.0 C 220kPa 6:31
136.4 C 221kPa 7:01
136.0 C 221kPa 7:31
136.4 C 222kPa 8:01
136.4 C 221kPa 8:31
136.4 C 221kPa 8:31
VENTING 8:32
AIR DRYING 9:28
CYCLE COMPLETE 25:16

Digital Signature #
E9A75A7BA4EE3F54
