

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3672

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2024 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 82.9 F 0.1  
4:00 77.4 F 0.6  
6:00 112.2 F 2.6  
8:00 145.2 F 5.2  
10:00 166.9 F 8.6  
12:00 185.7 F 12.8  
14:00 201.4 F 4.9  
16:00 216.6 F 8.5  
18:00 229.3 F 12.4  
20:00 230.1 F 8.3  
22:00 241.4 F 13.3  
24:00 240.2 F 11.3  
26:00 252.3 F 17.2  
28:00 262.5 F 23.4  
30:00 272.5 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.4 F 30.0  
2:00 271.5 F 29.2  
3:00 272.1 F 30.0  
4:00 272.1 F 29.7

Min 271.0 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 228.6 F 0.0  
10:00 191.0 F 0.0  
15:00 181.3 F 0.0  
20:00 177.4 F 0.0  
25:00 174.9 F 0.0  
30:00 173.5 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 30:28  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:08:20  
CYCLE COMPLETE

06/06/2024 09:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_