
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004948
15:32 10/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:00
136.0 C 220kPa 4:00
136.0 C 223kPa 4:30
136.3 C 224kPa 5:00
136.4 C 223kPa 5:30
136.1 C 225kPa 6:00
136.2 C 220kPa 6:30
136.3 C 224kPa 7:00
136.3 C 223kPa 7:30
136.3 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:22
CYCLE COMPLETE 9:45

Digital Signature #
B9BEF649F3212C34
