
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007816
9:09 03/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:33
PRESSURIZING 0:44
STERILIZING 1:48
136.0 C 220kPa 1:48
136.3 C 222kPa 2:18
136.4 C 221kPa 2:48
136.4 C 221kPa 3:18
136.4 C 222kPa 3:48
136.1 C 219kPa 4:18
136.1 C 220kPa 4:48
136.4 C 218kPa 5:18
136.4 C 218kPa 5:18
VENTING 5:18
AIR DRYING 6:08
CYCLE COMPLETE 6:20

Digital Signature #
EE30D839B9CC66CC
