
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000472
10:26 12/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:35
136.0 C 220kPa 4:35
136.4 C 232kPa 5:05
136.4 C 230kPa 5:35
136.1 C 229kPa 6:05
136.4 C 233kPa 6:35
136.4 C 233kPa 7:05
136.2 C 231kPa 7:35
136.4 C 233kPa 8:05
136.4 C 233kPa 8:05
VENTING 8:05
AIR DRYING 9:01
CYCLE COMPLETE 11:28

Digital Signature #
2E39637B05280E79
