

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3380

Name: RITTER2  
Sterilizer ID: V2313228  
02/13/2024 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.6 F 0.0  
2:00 133.6 F 0.7  
4:00 145.1 F 3.0  
6:00 177.1 F 7.0  
8:00 194.9 F 12.0  
10:00 213.2 F 6.2  
12:00 227.3 F 11.1  
14:00 230.4 F 7.6  
16:00 244.0 F 13.6  
18:00 242.8 F 12.3  
20:00 255.9 F 19.0  
22:00 267.3 F 26.2  
24:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.8 F 29.7  
2:00 272.2 F 30.0  
3:00 272.4 F 29.5  
4:00 272.3 F 30.1

Min 271.4 F 29.3  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
5:00 237.4 F 0.0  
10:00 191.6 F 0.0  
15:00 186.1 F 0.0  
20:00 183.7 F 0.0  
25:00 182.1 F 0.0  
30:00 180.6 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:05  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:02:15  
CYCLE COMPLETE

02/13/2024 12:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_