
Midmark M11 - v1.0.0
Total Cycles: 5082

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 74.0 F 0.0
4:00 86.5 F 0.0
6:00 126.8 F 0.1
8:00 154.1 F 0.9
10:00 178.1 F 0.8
12:00 201.4 F 2.6
14:00 211.3 F 5.8
16:00 223.6 F 9.0
18:00 231.6 F 12.8
20:00 232.3 F 7.8
22:00 243.8 F 12.8
24:00 240.6 F 10.3
26:00 236.4 F 8.3
28:00 251.1 F 15.2
30:00 262.6 F 21.8
32:00 272.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.7
1:00 272.6 F 28.1
2:00 272.0 F 28.0
3:00 272.3 F 27.9
4:00 272.3 F 28.2

Min 271.5 F 27.5
Max 272.9 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.6 F 0.0
5:00 204.1 F 0.0
10:00 225.4 F 0.0
15:00 233.0 F 0.0
20:00 230.1 F 0.0
25:00 233.5 F 0.0
30:00 227.8 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:17
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

11/13/2024 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____