
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001594
14:03 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:44
PRESSURIZING 6:27
STERILIZING 7:57
136.0 C 220kPa 7:57
135.8 C 217kPa 8:27
136.4 C 223kPa 8:57
136.1 C 220kPa 9:27
136.5 C 223kPa 9:57
135.8 C 220kPa 10:27
136.2 C 221kPa 10:57
136.1 C 219kPa 11:27
VENTING 11:27
AIR DRYING 12:27
CYCLE COMPLETE 17:11

Digital Signature #
D04C94C6CDBF041A
