
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002000
10:06 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:35
STERILIZING 3:36
136.0 C 220kPa 3:36
136.3 C 224kPa 4:06
136.4 C 224kPa 4:36
136.4 C 225kPa 5:06
136.2 C 222kPa 5:36
136.4 C 224kPa 6:06
136.3 C 224kPa 6:36
136.4 C 225kPa 7:06
136.4 C 225kPa 7:06
VENTING 7:06
AIR DRYING 8:00
CYCLE COMPLETE 11:15

Digital Signature #
8DBCFE837625E631
