
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001898
10:25 26/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:59
STERILIZING 3:05
136.5 C 227kPa 3:05
136.3 C 225kPa 3:35
136.4 C 225kPa 4:05
136.4 C 225kPa 4:35
136.4 C 226kPa 5:05
136.4 C 226kPa 5:35
136.1 C 221kPa 6:05
136.2 C 224kPa 6:35
136.2 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:25
CYCLE COMPLETE 33:44

Digital Signature #
849192FC0DCB5B0B
